











CooperVision changes the
way people see every day

	Silicone Hydrogel		Daily Disposables			Monthly Disposables			
	Biofinity*	Biofinity* Toric	Biomedics* 1 Day Extra	Biomedics* 1 Day Extra Toric	Proclear* 1 Day	Biomedics* 38	Biomedics* 55 Evolution	Biomedics* Toric	Proclear* Multifocal
									
Lens Material	comfilcon A	comfilcon A	ocufilcon D	ocufilcon D	omafilcon A	polymacon	ocufilcon D	ocufilcon D	omafilcon B
Water Content	48%	48%	55%	55%	60%	38%	55%	55%	62%
Base Curve	8.6	8.7	8.6	8.7	8.7	8.6	8.6	8.7	8.7
Diameter	14.0	14.5	14.2	14.5	14.2	14.0	14.2	14.5	14.4
Center thickness	0.08	0.11	0.07	0.11	0.09	0.04	0.07	0.11	0.16
Dk/t	160	116	27	17	36.6	23.5	27	18	21
Power Range	-0.50 to -6.00 (0.25 steps) -6.00 to -12.00 (0.50 steps)	Plano to -6.00 (0.25 steps) -6.00 to -10.00 (0.50 steps)	-0.50 to -6.00 (0.25 steps) -6.00 to -10.00 (0.50 steps)	Plano to -6.00 (0.25 steps) -6.00 to -10.00 (0.50 steps)	-0.50 to -6.00 (0.25 steps) -6.00 to -12.00 (0.50 steps)	-0.50 to -6.00 (0.25 steps) -6.00 to -10.00 (0.50 steps)	-0.50 to -6.00 (0.25 steps) -6.00 to -10.00 (0.50 steps)	Plano to -6.00 (0.25 steps) -6.00 to -9.00 (0.50 steps)	Plano to -6.00 (0.25 steps) -6.00 to -8.00 (0.50 steps)
Cylinder/ Addition		-0.75 -1.25 -1.75 -2.25		-0.75 -1.25 -1.75				-0.75 -1.25 -1.75 -2.25	+1.00 +1.50 +2.00 +2.50
Axis		10° to 180° (10° steps)		20°, 90°, 160° and 180° (Plano to -7.00) 90° and 180° (-7.50 to -10.00)				10° to 180° (10° steps)	
Design									D & N
Pack size	6	6	30	30	30	6	6	6	3



CooperVision®

CooperVision Contact Lens Malaysia Sdn Bhd

50-1-11 & 12, Level 1, Wisma UOA,

No. 50, Jalan Dungun, Damansara Heights, 50490 Kuala Lumpur, Malaysia

Tel: +603 2080 8488 Fax: +603 2093 7533

Website: www.coopervision.com.my • www.facebook.com/coopervisionmy

*Warning: UV-absorbing contact lenses are not substitutes for protective UV absorbing eyewear, such as UV absorbing goggles or sunglasses, as they do not completely cover the eye and surrounding area. Patients should continue to use UV-absorbing eyewear as directed.